

The reference for Impedance and Reflex Screening **Audiometer and Impedance Screener MI 26**



- Integrated audiometer, AC with 8 frequencies
- Automatic impedance test
- Fast, automatic test (only 3 seconds for one tympanogram)
- Ipsi- and contralateral (Option) reflex test with 4 frequencies
- Adjustable probe with integrated right / left selection button
- LCD screen displays tympanogram and test results immediately
- Graphical realtime reflex display
- Memory stores test results for both ears
- Fast, virtually silent printout of test results



The reference for Impedance and Reflex Screening Audiometer and Impedance Screener MI 26

Fast diagnosis of the middle ear

The MI 26 Impedance Screener with integrated audiometer performs fast impedance and ipsilateral reflex tests - a tympanogram takes only three seconds.

The fast, virtually silent printer offers a printout of the test result within only 4 seconds.

Contralateral reflex tests with MI 26C

Additionally, the model MI 26 C provides contralateral reflex measurements.

Versatile, adjustable probe

The probe tip can either be adjusted to straight, pen type (0° angle) operation or to 60° and 85° angle. An integrated switch allows to change the test ear directly at the probe. The integrated light changes to red (right ear) or blue (left ear) respectively.

Easy, comfortable to operate

The probe seal in the ear is easily checked with the built-in indicator light. The test starts immediately when a good seal is achieved. The tympanogram and test results are displayed in real time on the graphical LCD screen. The test results for both ears are stored and can be printed out together.

Automatic functions

The impedance test starts automatically. The intensity of the four frequencies of the following reflex test is increased automatically until a reflex is detected or the maximum is reached. The maximum level is adjustable to protect sensitive ears. An automatic printout of the test results can be selected.

Versatile use with integrated Audiometer

The versatile MI 26 offers in addition an integrated air conduction audiometer with 8 test frequencies. The audiograms for both ears are stored and can be printed out together.

Models: **MI 26**

MI 26 C with add. contralateral reflex test

Optional accessories / disposables

TDH 39 contralateral phone with headband		Part No.	4 6 8 2
Printer paper	1 roll	Part No.	7 0 5 0 7 8
1 Set of Eartips yellow (7 mm)	10 pcs.	Part No.	7 0 5 0 5 6
1 Set of Eartips green (9 mm)	10 pcs.	Part No.	7 0 5 0 5 7
1 Set of Eartips white (11mm)	10 pcs.	Part No.	7 0 5 0 5 8
1 Set of Eartips yellow (12,5mm)	10 pcs.	Part No.	7 0 5 0 5 9
1 Set of Eartips green (15mm)	10 pcs.	Part No.	7 0 5 0 6 0
1 Set of Eartips blue (18 mm)	10 pcs.	Part No.	7 0 5 0 6 1
Carrying case	1 pc.	Part No.	7 0 5 0 1 4



MI 26 with optional connector cover

Technical data

Tympanometry mode:

Probe frequency, intensity	226 Hz ± 1%, 85 dB SPL into 2 cm ³
Pressure range	+ 300 to - 600 daPa
Volume range	0,1 to 6,0 ml
Accuracy	± 5% or ± 10 daPa
Test time	< 3 seconds

Reflex Mode:

Test frequencies	500, 1000, 2000, 4000 Hz ± 2%
Test method	ipsilateral, contralateral (MI 26C)
Intensities ipsilateral	70 ... 105 dB HL
Intensities contralateral	70 ... 110 dB HL
Intensity setting	automatic or manual

Audiometer :

Test frequencies	0,25; 0,5; 1; 2; 3; 4; 6; 8 kHz
Dynamic range	0 to max. 90 dB _{HL}
Level steps	5 dB noise free
Headphones	Telephonics TDH 39

General:

Standard	IEC 601-1, IEC 645-5 class 2, IEC 645-1 class 4 medical directive 93/42/EEC
Test program	Reflex test selectable
Memory	Storage of test results for both ears
Menu	For adapting to individual needs (Pump speed, accuracy, reflex mode etc.)
Probe	Lightweight adjustable hand-held probe with built-in control light and switch
Printer	fast, virtually silent thermal printer
Print time	4 s / 12 s (Test results of both ears)
Display	Graphic LCD-Display, contrast adjustable
Power supply	Mains 96 ... 240 V~, 50/60 Hz, 25 VA
Dimension and weight	W x D x H: 39 x 29 x 11 cm / 2,6 kg
Standard access.MI26:	Probe, ear tip set, calibration cavity with probe holder, mains cable, roll of paper, headset TDH 39, patient response switch

Specifications are subject to change without notice



MAICO Diagnostic GmbH

Salzufer 13/14
D 10587 Berlin
Tel.: ++4930/70 71 46 50
Fax: ++4930/70 71 46 99

Email: export@maico-diagnostic.com
Internet: www.maico-diagnostic.com

850228/4 04/05